

Declaration of Conformity

BMT MESSTECHNIK GmbH herewith declares under sole responsibility the conformity of the products

OZONE ANALYZER BMT 965

OZONE ANALYZER BMT 964

OZONE MONITOR BMT 932

and all their modifications with the following regulations:

EMC Directive 2014/30/EU

Low-Voltage Directive 2014/35/EU

RoHS Directive 2011/65/EU and amendment (EU) 2015/863

These products contain less than 5 mg of Hg in their lamps. Exemption according to Annex III 4(a)-I: Mercury in low pressure non-phosphor coated discharge lamps, where the application requires the main range of the lamp spectral output to be in the ultraviolet spectrum: up to 15 mg mercury may be used per lamp.

Harmonized standards applied:



Safety: EN 61010-1: 2010
Safety requirements for electrical equipment for measurement, control, and laboratory use

EMC: EN 61326-1:2013
Electrical equipment for measurement, control and laboratory use – EMC requirements

Serial number: Each instrument has an identifiable serial number.

Year of construction: Each instrument has an identifiable year of construction.

Stahnsdorf, August 29, 2023



BMT MESSTECHNIK GMBH
Hamburger Str. 19
14532 Stahnsdorf
+49-3329-69677-0

Dr. Franz Wallner, President