

Certificate



Certificate no.

CU 72111657 01

License Holder:

BMT Messtechnik GmbH
Güterfelder Damm 87-91

14532 Stahnsdorf
Germany

Manufacturing Plant:

BMT Messtechnik GmbH
Güterfelder Damm 87-91

14532 Stahnsdorf
Germany

Test report no.: USA-DS 31083438 001

Client Reference: Ralph Christopher

Tested to:

UL 61010-1:2004 R10.08
CAN/CSA-C22.2 NO. 61010-1-04+G11 (R2009)

Certified Product: Ozone Monitor

License Fee - Units

Model Designation: BMT 932, BMT 932C

7

Rated Voltage:

1) AC 100-240V, 50/60Hz
2) optional: DC 18-36V SELV

Rated Power:


1) 35VA, 2) 30VA

Protection Class: I

Appendix: 1, 1-3

7

Licensed Test mark:



Signature


Dipl.-Ing. J. Engels
QA Certification Officer

Date of Issue

(day/mo/yr)
30/05/2011