

Certificate



Certificate no.

CU 72091486 01

License Holder:

BMT Messtechnik GmbH
Argentinische Allee 32a

14163 Berlin
Germany

Manufacturing Plant:

BMT Messtechnik GmbH
Güterfelder Damm 87-91

14532 Stahnsdorf
Germany

Test report no.: USA-DS 30981914 001

Client Reference: Ralph Christopher

Tested to: UL 61010-1:2004
CAN/CSA-C22.2 NO. 61010-1-04

Certified Product: Ozone Gas Analyzer

License Fee - Units

Model Designation: BMT 964 OZONE IN OFF-GAS ANALYZER

7

Rated Voltage: AC 100-240V, 50/60Hz
Rated Power: 120VA
Protection Class: I

7

Appendix: 1, 1-2

Licensed Test mark:

Signature

Date of Issue
(day/mo/yr)
30/06/2009



A handwritten signature in blue ink that reads 'Martin Glagla'.

Dipl.-Ing. M. Glagla
QA Certification Officer