

# Certificate



Certificate no.

CU 72181495 01

**License Holder:**

BMT Messtechnik GmbH  
Güterfelder Damm 87-91  
14532 Stahnsdorf  
Germany

**Manufacturing Plant:**

BMT Messtechnik GmbH  
Güterfelder Damm 87-91  
14532 Stahnsdorf  
Germany

Test report no.: USA- 31882167 001

Client Reference: Mr. R. Christopher

Tested to:

UL 61010-1:2012 R4.16

CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2

Certified Product: Ozone analyzer

License Fee - Units

Model Designation: 1) BMT 965 C, 2) BMT 965 OG

7

Rated Voltage: 1)+2) AC 100-240V, 50/60Hz

Rated Power: 1) 35VA, 2) 200VA


Protection Class: I

Degree of Protection: IP65

Appendix: 1, 1-6

7

Licensed Test mark:



Date of Issue

(day/mo/yr)

13/12/2018