## BMT 964 LINK Software Set-up

The LINK software is included with the 964 User Manual or is available for download at www.bmt-berlin.de/software

\*Connect the instrument to a Windows based computer using USB/Serial Adapter & Serial Cable provided with instrument (see below)







(for Standard Models)

When the LINK software is opened, it will try to connect to the BMT 964 series instrument. Once connected, the communication indicator will spin as the software receives data.

- Note: If software does not communicate with Analyzer, click the Options tab on the MAIN window and select COM Settings to select a different COM port (you may need to try several different COMM ports).
- Next: Once software is connected, on the MAIN window, click the View tab and select Diagnostics. This will open the Diagnostics window as shown below.

	Main	Wind	ow	
	🗛 BMT 964 Link #100742418	_		$\times$
	File View Options Help 🗃 🖫    🚭			
	Concentration: 0.1 g/Nm3 Pressure: 1.000 bar abs Cuvette status: clean (0.0%)			
	Range: 200.0 g/Nm3 Pressure Range: 1.15 bar abs Errors: none			
			E	at
Con	nmunication Indicator	•		

COMBI

Paran	neters Windo
💁 BMT 964 Link Parameters: 🛛 —	
Units Alarms I/O Gas Time	Other
6/17/2019 V	Jpdate 964
3:29:12 PM 📫 U	Jpdate 964
Date format: DD.MM.YY -	
PC Time: 6/17/2019, 3:29:45 PM Syr	nchronize 964
Date and Time are used for logging purp	oses, only.
Print Report	Close

When the Diagnostics window
opens and data populates the
fields, press the Print button
at the bottom (as shown).

When the Print window appears, select PDF as the printer, and Desktop as the file location. The BMT964 Link Diagnostics file should appear on your desktop.

This report includes current circuit performance indicators and Error and Event Logs which provide some historical data (see example on next page).

Please email this file to OSTI for analysis.

🗛 BMT 964 Link Diagnostics: #100742418		? X
<b>BMT 964</b> Hours running: 42753 Firmware: V2.01 <i>ran data vallage ran attict atticed</i> Ref. channet: 14599 3.665 V 49 0.012 V Meas. channet: 13750 3.453 V 47 0.012 V	Bit Var A: 3 ✓ calibrated ✓ VENT ─ PRESS ─ PRESS with 2 MPX	SysFlagA 1024 DispTime SetPara SecTick ScrollFlag ErterFlag BeepFast HiAlarm V UMode CMode DispBink DispBink DispBink DispBink AnalogSim ExternZero NoDisp RemDisp RemDisp RemDisp Bat 7 Bat 8 Bit 9 Bat 10 Bat 11 Bat 12 Bat 15 Bat 15 B
Lamp:         7579         2.635 V           Temperature:         1458         0.366 V         Initial Meas::         13666           Ameridade         vollaget         calive         gain           Vent Sensor:         10711         3.735 V         1.049000           Press sensor:         1209         0.304 V         0.000 V         1.000000           2nd MPX:         9         0.033 V         LDuiGP3           MPX source:         14272         4.962 V         0.000348           MPX true src:         4.961 V         100         3720.0           Gain:         87.154         SysFlag A:         0000 0100         0000 1000           Zero Factor:         0.942380         SysFlag B:         0000 1000         0000 1000	<ul> <li>AQ</li> <li>MPX4250</li> <li>590</li> <li>overpress. occured</li> <li>Bit Var B: 608</li> <li>date format m/d/y</li> <li>low alarm enabled</li> <li>high alarm enabled</li> <li>low alarm latching</li> <li>righ alarm latching</li> <li>ir alarm bacp on</li> <li>RS232 timed</li> <li>cuvete dirty warning</li> </ul>	
False Light: 0.1390 %       Bit Var A:       0000 0011         AD Factor 1.00429       Bit Var B:       0010 0110 0000         Binary input: 0000       Port 3: 255       Port 4: 253       Port 5: 255         Zero in       Image: Science of the second seco	Cuvette dity error     alarm relays closing     Errors & Status: 0     Iamp low earning     Iamp low error     Iamp off error     dity warning     dity error     overpressure error     overrange error     EEProm error     currently zeroing     warning up     Iamp high error	
Print to (Save As) .pdf file	•	

**Diagnostics** Window

Main LINK Window
💁 BMT 964 Link #100742418 — 🗆 🗙
File View Options Help
Concentration: 0.1 g/Nm3 Pressure: 1.000 bar abs Cuvette status: clean (0.0%)
Range: 200.0 g/Nm3 Pressure Range: 1.15 bar abs Errors: none
Exit
Communication Indicator
COM
Parameters Window
😋 BMT 964 Link Parameters: — 🗆 🗙
Units Alarms I/O Gas Time Other
6/17/2019 Vpdate 964

Update 964

Close

\*

-PC Time: 6/17/2019, 3:29:45 PM Synchronize 964 Date and Time are used for logging purposes, only.

Print Report

3:29:12 PM

Date format: DD.MM.YY

## Diagnostics Report - Email a copy to OSTI

